



Project
funded by the
EUROPEAN UNION



**ENPI
CBCMED**
EUROPEAN COOPERATION
IN THE MEDITERRANEAN



Errata corrige

Page 3 of the Special Conditions to the contract

P.11 CURRENT METER, acoustic, single-point, for measuring current, with a measuring range for speed from 0 to 300 m/sec, resolution 0.1 mm/sec, accuracy ± 0.15 cm/sec; measuring range for direction 0-360°, resolution 0.01°, accuracy $\pm 8^\circ$ including wave meter with measuring range from 0 a 400 kPa (58 psia), resolution $< 0.0001\%$ FSO, accuracy $\pm 0.02\%$ FSO. 15 POINTS

It is amended as follows

P.11 CURRENT METER, acoustic, single-point, for measuring current, with a measuring range for speed from 0 to 300 cm/sec, resolution 0.1 mm/sec, accuracy ± 0.15 cm/sec; measuring range for direction 0-360°, resolution 0.01°, accuracy $\pm 8^\circ$ including wave meter with measuring range from 0 a 400 kPa (58 psia), resolution $< 0.0001\%$ FSO, accuracy $\pm 0.02\%$ FSO. 15 POINTS

Page 7 of the Special Conditions to the contract

P.12 CURRENT METER acoustic single-point for current measurements with a measuring range for speed from 0 to 300 m/sec, resolution 0.1 mm/sec, accuracy ± 0.15 cm/sec; measuring range for direction 0-360°, resolution 0.01°, accuracy $\pm 8^\circ$ including wave meter with measuring range from 0 to 400 kPa (58 psia), resolution $< 0.0001\%$ FSO, accuracy $\pm 0.02\%$ FSO. 15 POINTS

It is amended as follows

P.12 CURRENT METER acoustic single-point for current measurements with a measuring range for speed from 0 to 300 cm/sec, resolution 0.1 mm/sec, accuracy ± 0.15 cm/sec; measuring range for direction 0-360°, resolution 0.01°, accuracy $\pm 8^\circ$ including wave meter with measuring range from 0 to 400 kPa (58 psia), resolution $< 0.0001\%$ FSO, accuracy $\pm 0.02\%$ FSO. 15 POINTS

Page 3 of the Tender regulations:

III.2.3 No. 1 of the notice: performed by the competitor in the period 2009/2014

It is amended as follows

III.2.3 No. 1 of the notice: performed by the competitor in the last three years preceding the date of publication of the notice

Firmato
Il Direttore Generale
Dr. Vincenzo Saggiomo